

CERTIFICATE OF ANALYSIS

Product : **MICRO AGAR**
 Product number : M1002
 Batch number : 015843
 Date of production : 18-11-2020
 Date of expiry : 30-11-2024

Packaging and storage : in a well-closed container, under dry and cool conditions


	SPECIFICATION	RESULT	UNITS	REMARKS
Appearance	powder	complies		
Colour	cream-white	complies		
Sieve Analysis	≤ 5	< 5	%	ASTM, on mesh 80
Gel strength (Nikkan)	900 – 1050	955	g/cm ²	after autoclaving
Loss on drying	≤ 10	7.08	%	
Total ashes	≤ 5	3.18	%	
pH	6 – 7.5	6.11		1.5% solution after autoclaving
Nephelos	≤ 10	4.14	NTU	after autoclaving
Melting point	> 85	91	°C	
Gelling point	34 – 36	35	°C	
Calcium	≤ 2000	1849	ppm	
Magnesium	≤ 1000	668	ppm	
Toxic substances	absence	absence		

This certificate of analysis is conform the results of the manufacturer.

Conclusion: approved
 Released by:

drs. M. van Wissen, pharmacist
 Department of Quality Control Duchefa B.V.

Date: 06-01-2021



The buyer is responsible to test each batch to ensure the product is suitable for their specific purpose.

Duchefa Biochemie B.V.